


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AV14-MPT-P*

S.N.: *6655*

Date: *MAY 13 1993*



*50.66 ATTEN*

*32 V/DIV*

*PRF = 1.0 MHz*

- a) Output signal amplitude:  
*≥ +30 VOLT*
- b) Pulse width: *TO 50*  
*≤ 1.0 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 VOLT*

*1.25 mA (MAX)*

*[Signature]*