


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-A1-PW-05-N-PS

S.N.: 6657

Date: APRIL 23 1993



$R_L = 50 \Omega$
PRF = 5.0 kHz

- a) Output signal amplitude:
0 to 200 V
- b) Pulse width:
50 ns to 1.0 μ s
- c) Rise time: (0.5% max duty cycle)
 ≤ 10 ns
- d) Fall time:
 ≤ 10 ns
- e) PRF: 0 to 100 kHz
0.5% max duty cycle
- f) Jitter, stability: cycle.
OK
- g) Prime power:
120/240 V
50-60 Hz