


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-1A-1-PW-05-N-P5*

S.N.: *6658*

Date: *MARIL 23 1993*



a) Output signal amplitude:

0 TO -200 V

b) Pulse width:

50 NS TO 1.0 μ S

c) Rise time: *0.5% MAX*
DUTY CYCLE

$\leq 10 NS$

d) Fall time:

$\leq 10 NS$

e) PRF:

0 TO 100 KHz

(0.5% MAX DUTY

f) Jitter, stability: *CYCLE*

OK

g) Prime power:

120/240 V

50-60 Hz

$R_L = 50 \Omega$

PRF = 5.0 KHz