


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVD-6C-C-PN-MA-SU1*


S.N.: *6664*

Date: *MAY 12 1993*



*P<sub>out</sub> R<sub>L</sub> = 50 Ω*

- a) Output signal amplitude:  
*0 TO ± 5 AMPS  
(0 TO ± 220V TO 40-50 Ω)*
- b) Pulse width:  
*0.1 TO 5.0 μS*
- c) Rise time: *(1% MAX DUTY CYCLE)*  
*≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF: *0 TO 10 KHz  
(1% MAX DUTY CYCLE)*
- f) Jitter, stability: *(CYCLE)*  
*OK*



*N<sub>out</sub> R<sub>L</sub> = 50 Ω*

- g) Prime power:  
*120/240 V  
50 to 1 kHz*

*[Signature]*