


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-6C-C-D-F1-EA*

S.N.: *6666*

Date: *JUNE 28 93*



a) Output signal amplitude:  
*0 TO + 5 AMPS*

b) Pulse width:  
*0.1 TO 5.0  $\mu$ S*



c) Rise time:  
*(10% MAX DUTY CYCLE)  
 $\leq 2 NS$*

d) Fall time:  
 *$\leq 2 NS$*

e) PRF: *0 TO 10 KHz*  
*(10% MAX DUTY CYCLE)*

f) Jitter, stability:  
*OK*

g) Prime power:  
*120/200 V*  
*50-60 Hz*



*5 AMP/DIV*  
*PRF = 5 KHz.*

