


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV6-4B-C-P-PN-ESMA*

S.N.: *6669*

Date: *JUNE 4 1993*



70 dB ATTEN
= 320 VOLTS/DIV
PRF = 10 KHz

- a) Output signal amplitude:
0 to ± 750 VOLTS
- b) Pulse width: *70 50 nS*
6 to 10 nS
- c) Rise time:
2 - 3 nS
- d) Fall time:
2 - 3 nS
- e) PRF: *1 Hz to*
10 KHz
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 - 60 Hz

LL