


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-6B-C-PN-M

S.N.: 6671

Date: JUNE 11 1993



- a) Output signal amplitude:
0 TO ± 2 AMPS
- b) Pulse width:
0.1 TO 5.0 MS
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF: 0 TO 10 KHz.
5% MAX OUTF VCO
- f) Jitter, stability:

OK

- g) Prime power:
120 / 240 V
50 - 60 Hz.

