


PULSE GENERATOR

PERFORMANCE CHECK

Model: *APR-3-PW-C-P-AT-PWT*


S.N.: *6672*

Date: *MAY 14 1993*



- a) Output signal amplitude:  
*0 to +200 VOLTS*
- b) Pulse width:  
*0.1 TO 100 NS*  
*(0.5% MAX DUTY CYCLE)*
- c) Rise time:  
*≤ 10 NS*
- d) Fall time:  
*≤ 10 NS*
- e) PRF: *1 KHz TO 10 KHz*  
*(0.5% MAX DUTY CYCLE)*
- f) Jitter, stability:  
*OK*

Ⓐ *R<sub>L</sub> = 50Ω, PRF = 1 KHz*



- g) Prime power:  
*120/240 VOLTS,*  
*50 - 60 Hz.*

*[Signature]*

Ⓑ *R<sub>L</sub> = 50Ω, PRF = 50 Hz*