


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV-143-02CB*

S.N.: *6673*

Date: *MAY 5 1993*



- a) Output signal amplitude:
+32V TO 5V
- b) Pulse width:
2 μ s
- c) Rise time:
 ≤ 200 ns
- d) Fall time:
 ≤ 200 ns
- e) PRF:
0 TO 13 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*+24 VDC
100 mA.*

*R_L = 50 Ω
PRF = 13 KHz*

*WILL WITHSTAND
OPEN & SHORT Ckt
LOAD COMPAN.*

CS