


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVO-601-C-P-451*

S.N.: *6676*


Date: *AUG 6 1993*



- a) Output signal amplitude:
0 TO +3.0 AMPS
- b) Pulse width:
50 NS TO 200 US
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK

*1.0 AMP / DIV
100 NS / DIV
PRF = 2 KHz*

- g) Prime power:
*120/240 V
50 - 60 Hz*



[Signature]