


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-C-N*  
S.N.: *6678*  
Date: *JUNE 8 1993*



- a) Output signal amplitude: *0 TO -30 V*
- b) Pulse width: *(TO 50ns)*  
*≤ 1ns*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*40 dB ATTEN*  
*10 V/DIV / DIV*  
*PRF = 100 kHz*

*120/240 V*  
*50-60 Hz*  
*[Signature]*