



PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-7A.*
S.N.: *6680*
Date: *JUNE 14 1993*



*P_{out} : R_L = 2.0 Ω
PRF = 500 Hz*

- a) Output signal amplitude:
0 to ± 20 VOLT?
- b) Pulse width: *TO R_L ≥ 2 Ω
100 NS TO 20 US
(49. MAX DUTY CYCLE)*
- c) Rise time:
≤ 30 NS
- d) Fall time:
≤ 30 NS
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability:
*(49. MAX DUTY CYCLE)
OK*



*N_{out} : R_L = 2.0 Ω
PRF = 5 KHz*

- g) Prime power:
*120 / 240 VOLT?
50 160 Hz.*