


PULSE GENERATOR

PERFORMANCE CHECK


Model: AV4-S-1-C-N

S.N.: 6687

Date: JUNE 23 1993



(A) 40dB ATTEN  
- 10 VOLT/10 DIV  
PRF = 100 KHz



(B) AS (A) BUT  
50 PS/PW  
(10 GHz BW SCOPE  
& ATTEN, (ANSWER))

- a) Output signal amplitude:  
0 TO 10 V
- b) Pulse width:  
≤ 130 PS
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power:  
120/240 V  
50-60 Hz  
44