


PULSE GENERATOR

PERFORMANCE CHECK

Model: ANH-MVI-71-N

S.N.: 6691

Date: JUN 29 1993



60 db ATTEN

100 VOLTS/DIV

PRF = 100 KHz

a) Output signal amplitude:
0 TO -100 VOLTS

b) Pulse width: (TO 50N)

≤ 0.6 NS

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 100 KHz

f) Jitter, stability:

OK

g) Prime power:

+15 VOLTS

150 mA (MAX)

[Signature]