


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-62-PS-P-TIA*

S.N.: *6702*

Date: *AUG 5 1993*



*100 VOLTS/DIV  
50 ns/DIV  
 $R_L = 1 \text{ M}\Omega$*

- a) Output signal amplitude:  
*0 TO +800 VOLTS*
- b) Pulse width:  *$R_L \geq 1 \text{ M}\Omega$   
10 NS TO 1 MS*
- c) Rise time:  
 *$\leq 50 \text{ NS}$*
- d) Fall time:  
 *$\leq 50 \text{ NS}$*
- e) PRF: *0 TO 10 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120/240 V  
50-60 Hz*  
*[Signature]*