


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-52-PS-P-PWT*

S.N.: *6703*

Date: *SEPT 7 '93*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width: *(TO 50ns)*
50 TO 500 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF: *0 TO 10 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN

100 VOLTS/DIV

PRF = 10 KHz

120/240 V
50 - 60 Hz
✓