


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVH-5-1-P-BLD

S.N.: 6708

Date: AUG 20 1993



- a) Output signal amplitude:
 $\approx +12 \text{ VOLT}$
- b) Pulse width:
 $\leq 100 \text{ PS}$
- c) Rise time: (FWHM)
 NA
- d) Fall time:
 NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability:
 OK
- g) Prime power:

40 dB ATTEN
10 VOLT/DIV
TOP: 500 PS/DIV
BOT: 50 PS/DIV

+15 VOLT DC
150 mA (MAX)