


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MV-155D-N-LHB-C*

S.N.: *6711*

Date: *JUL 7 16 93*




- a) Output signal amplitude:
0 TO -5.2 Amp
- b) Pulse width: *(0 TO 3 V)*
1 TO 10 MS (OR DC)
- c) Rise time:
≤ 10 US
- d) Fall time:
≤ 10 US
- e) PRF:
100 TO 500 Hz
(INT)
- f) Jitter, stability:
OK

TOP 2 VOLTS/DIV
 $R_L = 0.4 \Omega$
 2 MS/DIV

BOT 2 VOLTS/DIV
 $R_L = 0.4 \Omega$
 5 US/DIV
 (RISE TIME)

- g) Prime power:
-10 VOLTS, 6.0 Amp
MAX.
+10 VOLTS, 200 mA
MAX.



[Handwritten signature]