


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AMMP-4-C-PN*

S.N.: *6714*

Date: *AUG 16 1993*



(A) P_{out} 40 db attenu
10 VOLT/DIV
PRF = 1.0 MHz



(B) N_{out} $R_L = 50\Omega$
PRF \approx 5 kHz

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
10 NS TO 10 MS
- c) Rise time: *(10% MAX DUTY CYCLE)*
 ≤ 200 PS
- d) Fall time:
 ≤ 200 PS
- e) PRF: *0 TO 1.0 MHz*
- f) Jitter, stability: *(10% MAX DUTY CYCLE)*
OK
- g) Prime power:
170 / 240 V
50 - 60 Hz.

[Signature]