


PULSE GENERATOR

PERFORMANCE CHECK

Model: *MR-61-P5-F*

S.N.: *6720*

Date: *JULY 29 1993*



R_L = 10K.

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width: *(R_L = 10K)*
100 NS TO 1.0 μS
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 20 NS
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:
OK
- g) Prime power:

120 / 240 V
50 - 60 Hz

[Signature]