


PULSE GENERATOR

PERFORMANCE CHECK

Model: MV-1010-C-R5


S.N.: 6724

Date: APR 25 1993



- a) Output signal amplitude:
0 TO ± 100 VOLTS
- b) Pulse width: TO $R_L \geq 50 \Omega$
50 NS TO 10 MS
- c) Rise time:
 ≤ 15 NS
- d) Fall time:
 ≤ 15 NS
- e) PRF: 10 ITR TO 10 MHz
(10% MAX DUTY)
- f) Jitter, stability: CYCLE
- g) Prime power: $120/2000$ V
50-60 Hz

① P_{out} $R_L = 50 \Omega$
PRF = 100 KHz
(100 V RANGE)



② W_{out} $R_L = 50 \Omega$
PRF = 100 ITR
100 V RANGE


[Signature]

PULSE GENERATOR
PERFORMANCE CHECK

Model:

S.N.: 6724 CONT

Date:



a) Output signal amplitude:

b) Pulse width:

c) Rise time:


d) Fall time:

e) PRF:

f) Jitter, stability:

③ Load $R_L = 50\Omega$
PRF = 10 KHz
30 V RANGE
DOUBLE PULSE MODE

g) Prime power:



← AX-UMRC4 output
TO 3Ω. 100 VSB
IN FROM AV-1010L.

④