


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-3-PWC-AT-DWT-T*

S.N.: *6727*

Date: *JULY 23 1993*



- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
0.1 TO 100 μS
- c) Rise time: *(0.5% MAX DUTY CYC)*
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 10 KHZ
(0.5% MAX DUTY CYCLES)
- f) Jitter, stability:
OK
- g) Prime power:

R_i = 50Ω
PRF = 1.0 KHZ

120/240 V
50-60 HZ
[Signature]