


PULSE GENERATOR

PERFORMANCE CHECK

Model: *M-147-TR-A1*

S.N.: *6749*

Date: *OCT 11 1993*



- a) Output signal amplitude:  
*± 10 VOLTS (50n)*
- b) Pulse width:  
*50NS TO DC*
- c) Rise time:  
*≤ 2 NS*
- d) Fall time:  
*≤ 2 NS*
- e) ~~PRF:~~  
*OVERSHOOT: ≤ 5%*
- f) Jitter, stability:  
*OK*
- g) Prime power: *± 15 VOLTS (± 0.2V, 300 mA (MAX))*

