


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR14-3-PS-PN-PWT-AT


S.N.: 6760

Date: APR 31 1993



Pos 500 V/DIV  
 $R_L = 10K$

- a) Output signal amplitude:  
0 TO  $\pm 3000$  VOLTS
- b) Pulse width:  $R_L \geq 10K$   
700 NS TO 2.5  $\mu$ S
- c) Rise time:  
 $\leq 100$  NS
- d) Fall time:  
 $\leq 100$  NS
- e) PRF:  
0 TO 1 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/240 V  
50-60 Hz



Neg 500 V/DIV  
 $R_L = 10K$

