


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVR-A1-PW-C-P*

S.N.: *6770*

Date: *AUG 9 1993*



$R_i = 50 \Omega$
 $PRF = 10 \text{ kHz}$

- a) Output signal amplitude:
0 TO +200 VOLTS
- b) Pulse width:
50 NS TO 1.0 μ S
- c) Rise time:
(0.5% MAX DUTY CYCLE)
 $\leq 10 \text{ NS}$
- d) Fall time:
 $\leq 10 \text{ NS}$
- e) PRF: *0 TO 100 kHz*
(0.5% MAX DUTY CYCLE)
- f) Jitter, stability: *(CYCLE)*
- g) Prime power: *OK*

120 / 200 VOLTS
50 - 60 kHz