


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-E3.-C-P-05

S.N.: 6772

Date: SEPT 10 1993



50 dB ATTEN

32 VOLT/DIV

PRF = 20 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
(TO 50V)
 - b) Pulse width:
20 TO 200 NS
 - c) Rise time:
≤ 0.5 NS
 - d) Fall time:
≤ 1.0 NS
 - e) PRF:
0 TO 20 KHz
 - f) Jitter, stability:
OK
 - g) Prime power:
10/200 ✓
50 - 60 Hz
- A-