


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-1V-1-ECL-N*

S.N.: *6777*

Date: *SEPT 17 1993*



- a) Output signal amplitude:
0 TO - 10 VOLTS
- b) Pulse width:
0.2 TO 4.0 N
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
400 mA.

40 db ATTEN
10 VOLTS/DIV
PRF ≈ 300 kHz

h) INPUT TRAIL:
ECL

INPUT DOES NOT
INCLUDE 50Ω OR
50Ω -2V TERMINATION