


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMC-E4-C-P-0514-M5LB*

S.N.: *6780*

Date: *OCT 19 1993*



60 dB ATTEN

*100 VOLTS/DIV
(RANGE B)*

PRF = 50 KHz.

R_L = 50 Ω.

- a) Output signal amplitude: *0 TO +100 VOLTS
(TO 50 Ω)*
- b) Pulse width: *10 NS TO 200 NS
(0.5% MAX DUTY CYCLE)*
- c) Rise time: *≤ 1 NS*
- d) Fall time: *≤ 1 NS*
- e) PRF: *0 TO 100 KHz
(0.5% MAX DUTY CYCLE)*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
57-60 Hz*

AF