


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVB1-3-P5C-500
- 546
S.N.: 6781, 82, 83, 84, 85, 86 - 625
- 690
Date: SEPT 28 '93 - 770
- 862



60 dB ATTEN
100 VOLTS/DIV
PRF = 100 KHz
(- 862 UNIT)

a) Output signal amplitude:
→ +100 VOLTS

b) Pulse width:

c) Rise time: NA

d) Fall time: NA

e) PRF: 0 TO 160 KHz

f) Jitter, stability:
OK

g) Prime power:
+15 VOLTS DC
250 mA MAX