


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVMM-1-C-D-M-OT-PN

S.N.: 6788

Date: NOV 1 1993



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 60 nS
- c) Rise time:
 ≤ 100 pS
- d) Fall time:
 ≤ 135 pS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK

40 dB ATTEN
10 VOLTS/DIV
PRF = 3.0 MHz

- g) Prime power:
140/200 V
52 60 Hz