


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-2A-C-P-LA1 (+ AVX-S3A-MF-LA1A)

S.N.: 6799 (+ 6803)

Date: SEPT 22/93



- a) Output signal amplitude:
0 TO + 5.0 AMP
- b) Pulse width:
5 TO 50 NS
- c) Rise time:
 ≤ 2 NS
- d) Fall time:
 ≤ 2 NS
- e) PRF:
10, 20, 1000 Hz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 Hz

M_I OUT TO 30db

$\therefore \approx 3.2$ VOLTS/DIV

$\therefore \approx 5.0$ AMP

1N459A DIODE USED

TO SIMULATE L. D.