


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-3-PW-PS-05-CA-N

S.N.: 6804

Date: SEPT 27 1993



50 VOLTS / DIV
 $R_L = 50 \Omega$
PRF = 1 KHz

- a) Output signal amplitude:
0 TO 200 V
- b) Pulse width:
0.1 TO 100 ns
(0.1% max duty cycle)
- c) Rise time:
 ≤ 10 ns
- d) Fall time:
 ≤ 10 ns
- e) PRF: 0 TO 1 KHz
(0.5% max duty cycle)
- f) Jitter, stability:
OK
- g) Prime power:
120/200 V
50-60 Hz