


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PII-V-142-CGA*

S.N.: *6812*

Date: *Nov 5 1993*



60 db ATTEN
∴ 100 VOLT/DIV
PRF ≈ 20 KHz

- a) Output signal amplitude:
0 TO +100 VOLTS
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
≤ 2 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VDC
250 mA (MAX)

[Signature]