


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-4B-CAN-01-M*

S.N.: *6817*

Date: *DEC 23 1993*



- a) Output signal amplitude:
0 to ± 850 VOLTS
- b) Pulse width:
≤ 8 nS
- c) Rise time:
≤ 3 nS
- d) Fall time:
≤ 3 nS
- e) PRF:
0 to 10 kHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN
≈ 100 VOLTS PIN
PRF = 5 kHz

140/200 V
50-60 Hz
[Signature]