


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV4-14V1-PN-55c2*

S.N.: *6840*

Date: *NOV 3 1993*



- a) Output signal amplitude:
± 230 uolts
- b) Pulse width:
2.15 ns
- c) Rise time:
≈ 1 ns
- d) Fall time:
≈ 14 ns
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:
- g) Prime power: *OK*
+15 uolts
400 mA
max.

ⓐ 60 db ATTEN
100 uolts/div
PRF ≈ 30 KHz.