


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMH-5-HV1-C-N*

S.N.: *6342*

Date: *NOV 29 1993*



- a) Output signal amplitude:
0 TO -40 VOLTS
- b) Pulse width:
≤ 350 PS (FWHM)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS

40 dB ATTEN

-10 VOLTS/DIV

PRF = 100 kHz

50-60 kHz

NA