


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AM-63-PS-PE22A2*

S.N.: *6863*

Date: *DEC 22 83*



a) Output signal amplitude:

*120 V*

b) Pulse width:

*T<sub>PF</sub> 120 + 150 ns*

c) Rise time:

*T<sub>RI</sub> 1 ns*

d) Fall time:

*T<sub>FL</sub> 2 ns*

e) PRF:

*5 to 10 kHz*


f) Jitter, stability:

*OK*


g) Prime power:

*120/200 V*  
*50-60 Hz*

*T<sub>PPA</sub>: 100 V*  
*5 kHz*



*T<sub>PPB</sub> 100 V*  
*5 kHz*



*20V*

*SAN B*