


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-V-C-P-M*
S.N.: *6865*
Date: *NOV 10 1993*



- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width:
2 TO 100 NS
- c) Rise time:
≤ 350 ps
- d) Fall time:
≤ 2 ns.
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK.
- g) Prime power:

TOP *50 db*
32 VOLTS/DIV
10 NS/DIV

130T *AS TOP BUT*
500 ps/DIV
(RISE TIME)

PRF = 10 KHZ

120/240 V
50-60 HZ.