


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVP-AV-1-PS-N-PW-S1*

S.N.: *6867*

Date: *FEB 16 1994*



40 db ATTEN
- 10 VOLTS/PIN
PRF = 100 KHz

- a) Output signal amplitude:
0 TO 710 VOLTS
- b) Pulse width:
0.2 TO 4.0 μs
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

90 TO 110 VOLTS
50 - 60 Hz.

VOLTAGE SELECTION
CRANK TO BE
IN 120 V POSITION
(AS SHIPPED).