


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AGR-3-PW-C-PSA P*


S.N.: *6868*

Date: *DEC 23 1993*



- a) Output signal amplitude:  
*0 TO + 200 VOLTS*
- b) Pulse width:  
*0.1 TO 100 NS*
- c) Rise time:  
*≤ 20 ns (CL = 0 pF)*
- d) Fall time:  
*≤ 1 ns (CL = 5000 pF)*
- e) PRF: " " "  
*0 TO 10 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 60 Hz*

① *R<sub>L</sub> = 10K PRF = 1 KHz*  
*C<sub>L</sub> = 0 pF*



② *R<sub>L</sub> = 10K*  
*C<sub>L</sub> = 4700 pF*  
*PRF = 1 KHz*