


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PH-5-C-P-M1*

S.N.: *6875*

Date: *FEB 22 1994*



- a) Output signal amplitude: *0 TO 118 V.*
- b) Pulse width: *≤ 200 pSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1.6 MHz*
- f) Jitter, stability: *OK*
- g) Prime power:

*40 db ATTEN*  
*10 VOLT/DIV*  
*PRF = 200 kHz*

*120/240 V*  
*50-60 Hz*