


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-P*

S.N.: *6885*

Date: *FEB 24 1994*



- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
≤ 4.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
250 mA (MAX)

60 dB ATTEN

100 VOLTS/DIV

PRF = 30 KHz

