


PULSE GENERATOR
PERFORMANCE CHECK

Model: *1720-53-C-P*

S.N.: *6888*

Date: *DEC 21 1993*



- a) Output signal amplitude:
0 TO + 48 AMP
- b) Pulse width: *(TD. 0.75 μ s)*
25 TO 500 ns
- c) Rise time:
 ≤ 10 ns
- d) Fall time:
 ≤ 10 ns
- e) PRF: *0 TO 1 KHz*
- f) Jitter, stability:
OK

$R_L = 0.7 \Omega$
 $\therefore \approx 68.3$ AMP/DIV
PRF = 1.0 KHz

g) Prime power:
120/240 VOLTS
50-60 KHz