


PULSE GENERATOR  
PERFORMANCE CHECK

Model: 'AVL-3AH-PS-P

S.N.: 6892

Date: MAR 14 1994



- a) Output signal amplitude:  
0 to +450 V
- b) Pulse width: (to 50 n)
- 5 to 100 nS
- c) Rise time:  
≤ 1 nS
- d) Fall time:  
≤ 4 nS
- e) PRF: 0 to 5 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

70 dB ATTEN  
3 320 VOLTS/DIV  
PRF = 2.0 kHz

120/240 V  
50-60 Hz  
*[Signature]*