


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-N-C
 S.N.: 6898
 Date: JAN 18 94



- a) Output signal amplitude: 0 TO -50 VOLT
- b) Pulse width: (70%)
2 TO 50 NS
- c) Rise time: ≤ 350 P
- d) Fall time: ≤ 2 NS
- e) PRF: 0 TO 20 KHZ
- f) Jitter, stability: OK

TOP 50dB ATTN: 32 VOLT/DIV
 10 NS/DIV

BOT 1 NS/DIV

PRF = 20 KHZ.

g) Prime power: 120/240 V
50-60 KHZ

[Handwritten signature]