


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APP-1-C-N*


S.N.: *6900*

Date: *MAR 22 94*



- a) Output signal amplitude:
0 TO -10 V
- b) Pulse width:
0.2 TO 100 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 Hz

① A RANGE
40 dB
-10 VOLTS/DIV
PRF = 100 KHz



② B RANGE
40 dB
-10 VOLTS/DIV
PRF = 100 KHz

M