


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PH-121-PN-55C1*

S.N.: *6906*

Date: *JAN 6 97*



- a) Output signal amplitude:
 $\pm 100 \text{ V}$
- b) Pulse width:
0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 kHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN

*+15V 250 mA
MAX*

100 VOLT/DIV

PRF = 50 kHz