


PULSE GENERATOR
PERFORMANCE CHECK

Model: *APR-3-PW-P-MITD*

S.N.: *6920*

Date: *MAR 1 94*



$R_L = 2 \Omega$

- a) Output signal amplitude:
0 to +30 VOLTS
- b) Pulse width: *$R_L = 2 \Omega$*
- c) Rise time: *$\leq 200 \text{ ns}$*
- d) Fall time:
 $\leq 200 \text{ ns}$
- e) PRF: *0.5% MAX
DUTY CYCLE.*
- f) Jitter, stability:
OK

g) Prime power:

- a) *+24 VDC
150 mA.*
- b) *0 to +32 VOLTS
1.0 AMP*