


PULSE GENERATOR
PERFORMANCE CHECK

Model: PI-V-HV2-COA-P-CA-6W

S.N.: 6933

Date: JAN 31 94



- a) Output signal amplitude:
0 TO +100 V
- b) Pulse width:
0 TO 20 NS
- c) Rise time:
 ≤ 2 NS
- d) Fall time:
6 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

60 dB ATTEN

100 VOLTS/DIV

PRF \approx 20 KHz

+15 VOLTS DC
250 mA MAX.