


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PRG-A 1-PW-C-N-FU*

S.N.: *6973*

Date: *MAY 31 1994*



R_c = 50 Ω
PRF ≈ 50 kHz

- a) Output signal amplitude: *0 TO -200 VOLTS*
- b) Pulse width: *(20 50 ns)*
- c) Rise time: *50 TO 100 ns*
- d) Fall time: *≤ 10 ns*
- e) PRF: *0 TO 50 kHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120 / 240 VOLTS*
50-60 kHz

[Signature]